

Product Name	CTNGlimy2mgTab5x2x15'sDomR				
Product Code	150098128	Version No.	NA	Std. Limit (%W/W)	NULL
Batch No.	Batch_5896	Equipment ID	kai_check	Friabilator Code	None
Prod. Version No	NA	Cubicle Name	IPQA-11-1330	Balance ID	3365220

Date	StartTime	Friability			End Time	Remark	Done By
		Initial Weight Before Test (g) A	Final Weight After Test (g) B	% Friability (A-B)/Ax100			
27.08.2024	12:39:09	79.3	78.3	1.261	00:00:00	Out Of Limit	Nick(22)

Printed By : Nick(22)

Date & Time : 27.08.2024 14:06:46