

LOSS ON DRYING (LOD) TEST RECORD
(GRANULATION)

Product / Method Name	DOLO		
Date	17.01.2025	Product Code	11
Batch No.	2123232	MPR No.	LOD
SFOID	3	Make	Mettler Toledo
Model	HX204	Instrument ID	PR703
Serial No.	C004959123	Repetition No.	3
End Time	17:09	Switch Off Mode	5:00 min
Total Time	5:00 min	Drying Program	Standard
Loss On Drying Range	(1.5-3.5) % w/w (Target of 2.5% w/w)	Dry Temperature (°C)	105

Test Results	
Initial Weight (g)	2.635
Dry Weight (g)	2.546
End Result (%MC)	3.500

Result: **Pass**

Performed by : Jaisukh BK
Date : 17.01.2025