

Print Date : 08.10.2025

Print Time : 10:31:13

Print By : Jillella Nagaraju

Product Name : Printing of Pantopra.Sod.DR Tab 40mg USA

DR.REDDY'S LABORATORIES LTD

FTO Unit-3

PRE965 - Batch Report Disintegration Time (mm:ss)

Batch Date : 08.10.2025

Batch No : AVG1

Product Code	200013010
Compression Equipment ID	PRE965
Type of Rotary	Double
Cubicle Name	CCOM
Disintegration tester ID	PRE2181
No of Stations	77

Batch Start Time	08.10.2025 10:26:52
Batch Stop Time	08.10.2025 10:30:40
Username	Jillella Nagaraju

S. No.	Start Time	Stop Time	Instrument ID	LHS Container No	RHS Container No	NMT(mm:ss)	Min	Max	Avg	Result	UserName
1	08.10.2025 10:27:42	08.10.2025 10:27:42	PRE2181	3412	NA	15:00	13:12	13:12	13:12	PASS	Initial User
2	08.10.2025 10:29:47	08.10.2025 10:29:47	PRE2181	424324	NA	15:00	10:10	10:10	10:10	PASS	Initial User

Avg Of Avg	00:11:41
------------	----------

Print Date : 08.10.2025
Print Time : 10:31:15
Print By : Jillella Nagaraju
Product Name : Printing of Pantopra.Sod.DR Tab 40mg USA

DR.REDDY'S LABORATORIES LTD
FTO Unit-3
PRE965 - Batch Report Disintegration Time (mm:ss)

Batch Date : 08.10.2025
Batch No : AVG1

