





BATCH REPORT OF FRIABILITY TEST
Area : Compression

Product Name	CTNGlimy2mgTab5x2x15'sDomR		
Product Code	150098128	Make	null
Batch No.	Batch_5896	Stage	null
Model No.	null	Std. Limit NMT(%W/W)	1.0

Batch Summary		
Minimum Friability(%w/w)	Maximum Friability(%w/w)	Average Friability(%w/w)
0.00	3.28	1.19



Detailed Report									
Test No.	Date	Time	Friabilator ID No.	Side	Initial Weight	Final Weight	Friability (%w/w)	Test Result	Performed By
1	2024-08-28	17:08:41.00 00000	MB-001	LEFT	0	0	0	Complies	Nick
2	2024-08-28	17:08:41.00 00000	MB-001	RIGHT	0	0	0	Not Complies	Nick

BATCH REPORT OF FRIABILITY TEST

Area : Compression

Test No.	Date	Time	Friabilator ID No.	Side	Initial Weight	Final Weight	Friability (%w/w)	Test Result	Performed By
3	2024-08-29	10:05:29.00 00000	MB-001	LEFT	28.0	27.9	0	Complies	Nick
4	2024-08-29	10:05:29.00 00000	MB-001	RIGHT	28.0	27.4	0	Not Complies	Nick
5	2024-08-29	10:59:03.00 00000	MB-001	LEFT	27.4	27.0	1.46	Not Complies	Nick
6	2024-08-29	10:59:03.00 00000	MB-001	RIGHT	27.4	26.5	3.28	Not Complies	Nick
7	2024-08-29	11:00:59.00 00000	MB-001	LEFT	27.4	27.0	1.46	Not Complies	Nick
8	2024-08-29	11:00:59.00 00000	MB-001	RIGHT	27.4	26.5	3.28	Not Complies	Nick