



# pH Calibration Report

Thermo Scientific (c) 2011

Versa Star Pro

Meter S/N V19961  
SW Rev 11.50

Channel 1  
Module S/n VA30652  
Method 101  
Calibration Time 28-07-25,16:48

-- pH Calibration Report --

Point 1  
pH 6.90  
mV 0.0  
Temperature 25.0 C (MAN)  
Calibration Type Manual

Average Slope 98.4 %  
Cal # 108

Meter ID : Test1  
Probe ID : PH@123

Tolerance : 90.00 % - 105.00 %  
Result : Pass  
Done By : Thermofisher