



pH Calibration Report

Thermo Scientific (c) 2011

Versa Star Pro

Meter S/N	V14010
SW Rev	11.50
Channel	1
Module S/n	VA18288
Method	100
Calibration Time	29-07-25,11:25

-- pH Calibration Report --

Point	1
pH	7.00
mV	0.0
Temperature	25.0 C (MAN)
Calibration Type	Manual
Average Slope	98.3 %
Cal #	105

Meter ID :	12345
Probe ID :	123PH

Tolerance :	90.00 % - 105.00 %
Result :	Pass
Done By :	User1004