



pH Calibration Report

Thermo Scientific (c) 2011

Versa Star Pro

Meter S/N V19961
SW Rev 11.50

Channel 1
Module S/n VA30652
Method 101
Calibration Time 29-07-25,10:29

-- pH Calibration Report --

Point 1
pH 7.00
mV 0.0
Temperature 25.0 C (MAN)
Calibration Type Manual

Average Slope 98.4 %
Cal # 109

Meter ID : Test1
Probe ID : PH@123

Tolerance : 90.00 % - 105.00 %
Result : Pass
Done By : User1002