

Conductivity Test Slip

Reprint

Thermo Scientific (c) 2011

Versa Star Pro

Meter S/N V18806
SW Rev 11.50

Channel - 1

Module S/n # VA30043
Mode Conductivity
Conductivity 98.25 uS/cm
Conductance 1056.489 uS
Temperature 24.7 C (ATC)
Temp. Coefficient Off
Cell Constant 0.0930 /cm
Method 100
CalLog # 0
Last Cal -----
Log # 27

Date & Time : 26-11-2024 & 12:56:37

Product Name : 100 microsemens

Sample ID : NA

AR No : NA

Batch No : NA

Product Type : Calibration Buffer

Meter ID : QC-CONMZ-002

Probe ID : ZO1-16440

Done By : Trident

Reviewed By : Trident

Comment : As part of OQ

Date & Time : 13-09-2024 & 13:01:46

Printed By : Thermofisher

Date & Time : 16-01-2025 & 11:28:02