

pH Calibration Report

Reprint

Thermo Scientific (c) 2011

Versa Star Pro

Meter S/N V122320
SW Rev 11.50

Channel 1
Module S/n VA36002
Method 100
Calibration Time 01-11-24,07:06

-- pH Calibration Report --

Point 1
pH 1.68
mV 304.1
Temperature 24.4 C (ATC)
Calibration Type Auto

Point 2
pH 4.01
mV 165.5
Temperature 24.3 C (ATC)
Calibration Type Auto

Point 3
pH 7.00
mV -8.1
Temperature 24.8 C (ATC)
Calibration Type Auto

Point 4
pH 10.01
mV -182.1
Temperature 25.0 C (ATC)
Calibration Type Auto

Point 5
pH 12.46
mV -328.9
Temperature 24.9 C (ATC)
Calibration Type Auto

Eo -7.8 mV
Average Slope 99.1 %
Cal # 102

Meter ID : QC-PHMTZ-007

Probe ID : CX1-15733

Tolerance : 90.00 % - 105.00 %

Result : Pass

Done By : B. Ramvikas

Reviewed By : S. Siddhartha Kumar

Comment : Reviewed

Date & Time : 01-11-2024 & 07:43:03

Printed By : Thermofisher

Date & Time : 16-01-2025 & 11:25:51